Certificate of Analysis

Product Name: **Bosphore Atypical CAP Panel Kit**

Catalog Number: **ABCAP1-2-3**

Lot Number: **SM6009**

Expiry Date: **04.2026**

Quality Control

Sensitivity and reproducibility of all the components included in Bosphore Atypical CAP Panel Kit were tested in the Real-Time PCR quality control experiment performed on 07.10.2024. Performance of the kit was verified with regard to the acceptance criteria stated in the user manual.

Shipping and Storage

Bosphore Atypical CAP Panel Kit is shipped on dry ice. It should be stored immediately upon receipt at -20 °C in a constant-temperature freezer and protected from light. When stored under the conditions and handled correctly, the products can be kept at least until the expiration date without showing any reduction in performance.

Certified by: ___Ozancan Ulaş_____

Inspection Date: 07.10.2024