FlashDx Shenzhen Inc.
Address: B6-4C, B5-A201, No.2533 Guanguang Road,
Fenghuang Community, Fenghuang Street, Guangming District
Shenzhen, Guangdong, P. R. China.
Phone: +86 (0)755-86965752

Certificate of Analysis

Report No. RQ250110001

Product Name	Respiratory Panel 1.1		
Catalog No.	RP-P10-10	**Specification**	10 tests/ box
Lot No.	S025F103		
Manufacture Date	2025-01-09	**Expiry Dated**	2026-01-08

QC Test Results:

Test Document	FlashDx Respiratory Panel 1.1 Test Quality Control Specification	
Test Sampling Amount	20 Tests	
Test Item	**Test Criteria**	**Test Result**
Appearance	All components of the cartridge are complete and intact. The package label shall be clear and no wear. The aluminum foil bag containing the cartridge is sealed without damage.	Pass
Compliance rate of positive reference	The testing results of positive references shall be all positive for the corresponding assay tests.	Pass
Compliance rate of negative reference	The testing results of negative references shall be negative for all assay tests. Internal control is valid, Ct≤35	Pass
Limit of Detection (LoD)	The testing results of LoD reference shall be positive for corresponding assay tests.	Pass
Overall Test Conclusion		**QC Pass**

FLASHDX SHENZHEN INC. 深圳闪量科技有限公司 品质部 QA Dept.

Tested by:	**Reviewed by:**	**Approved by:**
Ruina Li	Wanming Xian	Yongjun Yang
Date: 2025.01.10	**Date:** 2025.01.10	**Date:** 2025.1.10